



Operator's Manual

Power Analysis Software



Power Analysis Software Operator's Manual

© 2013 Teledyne LeCroy, Inc. All rights reserved.

Unauthorized duplication of Teledyne LeCroy documentation materials other than for internal sales and distribution purposes is strictly prohibited. However, clients are encouraged to distribute and duplicate Teledyne LeCroy documentation for their own internal educational purposes.

Power Analysis and Teledyne LeCroy are registered trademarks of Teledyne LeCroy, Inc. Windows is a registered trademark of Microsoft Corporation. Other product or brand names are trademarks or requested trademarks of their respective holders. Information in this publication supersedes all earlier versions. Specifications are subject to change without notice.

921326 Rev B

June 2013

Contents

- Power Analysis Software Overview** **2**
- Software Overview** **3**
- Required Equipment 3
- Method of Operation 4
- Power Analysis Software Dialogs 4
- Deskew Voltage and Current Channels 6
- Power Device Analysis** **9**
- Device Analysis Preliminary Setup 9
- Set Up Device Analysis Input Channels 11
- Identify Switching Zone 12
- Device Losses Test 13
- Device Safe Operating Area Test 15
- Device B-H Curve Test 16
- Device Dynamic On-resistance Test 17
- Device dv/dt Test 18
- Control Loop Analysis** **19**
- Control Loop Analysis Preliminary Setup 19
- Set Up Control Loop Analysis Source Channel 21
- Closed Loop Test 22
- Line Power Analysis** **23**
- Line Power Analysis Preliminary Setup 23
- Set Up Line Power Analysis Source Channels 23
- Line Power Test 25
- Line Harmonics Test 26
- Performance Analysis** **28**
- Performance Efficiency Test 28
- Performance Ripple Test 30

Power Analysis Software Overview

The Power Analyzer Software (PAS) option helps measure and analyze the operating characteristics of power-conversion devices and circuits for off-line, DC-DC and DC-AC power circuit designs. It provides automatic detection and measurement of turn-on and turn-off switching device losses, as well as conduction losses. Areas of power loss are clearly delineated by a color-coded waveform overlay, and tools are provided to reduce measurement errors. A streamlined user interface guides you through the various stages of analysis.

PAS consists of these major analysis areas:

- [Device Analysis](#) covers the measurement of switching device performance such as device instantaneous power, switching losses, safe operating area (SOA), saturation voltage, dynamic on-resistance, dv/dt or di/dt , and saturation hysteresis curves of magnetic devices. With PAS, these device measurements can be made either on a test stand or in circuit while the devices are operating in a power conversion system.
- [Control Loop Analysis](#) covers the acquisition and analysis of information contained in a power conversion circuit's modulated control signal. It analyzes modulation changes in pulse width (PWM), duty cycle, frequency, or period as the feedback loop responds to changes in line and/or load, as well as during start-up and shut-down.
- [Line Power Analysis](#) covers the measurement of line voltage and current applied to an off-line power conversion device. Real power, apparent power, power factor, crest factor, and line harmonics are measured. Analysis of line harmonic content is included to assist the design and evaluation engineer in designing for pre-compliance to EN 61000-3-2 requirements.
- Performance Analysis computes the efficiency of the power supply and adds the measurements to determine the amount of ripple a power source generates.

All Teledyne LeCroy [voltage and current probes](#) are integrated with the software, and measurements are automatic and precise. Documentation and Flashback to prior power-circuit analyses can be accomplished through LabNotebook.

Software Overview

Required Equipment

You will need this equipment to utilize the Power Analysis Software.

Oscilloscope

Power Analyzer Software option operates on any Teledyne LeCroy Windows-based oscilloscope. For analysis of phenomenon requiring the acquisition of many cycles, an oscilloscope with a minimum memory of 1 Mpt per channel is recommended. Steady-state analysis can be accomplished with shorter record lengths. Two acquisition channels are adequate for most measurements, but a four-channel oscilloscope is recommended if you would like to analyze multiple devices or use complex triggering. To see the small signal details hidden in large signals, such as a saturation voltage, or ripple transient analysis, a 12-bit oscilloscope is recommended.

Probes

Voltage Probes. A wide range of voltage probes are available and are integrated within this software. Proper selection of the following probes should match circuit details, including: single-ended passive, single-ended high voltage, differential high-voltage, active single and differential high-frequency, and 50 Ohm transmission line probe. Differential voltage probes will support your measurement environment including isolating your circuits from line power, measuring currents with current sensing resistors, and viewing switching transients in power supply ripple measurements.

Current Probes. The measurements described in Power Device Analysis require precision wide-bandwidth current probes with DC measurement capability. We recommend Teledyne LeCroy:

- **AP015** or **CP030** DC-to-50 MHz, 30-ampere current probe
- **CP031** DC-to-100 MHz, 30-ampere current probe
- **CP150** and **CP500** high-current probes

Other, higher current probes are also available from Teledyne LeCroy.

Teledyne LeCroy ProBus probes automatically use correct units and scaling for power measurements when used with the Power Analyzer software. When other probes are used, the Power Analyzer Software provides methods for entering the correct units and scaling.

When a channel is selected as the current input within the software, its units are automatically changed to Amperes. When a differential probe is used to measure the voltage across a shunt resistor, the Power Analyzer software will support the proper amps/div scaling when the resistor value is entered.

Differential Amplifier

Measuring high-side gate drive signals in an off-line application and capturing a device saturation voltage to measure conduction loss or $R_{ds(on)}$ are challenging to do. These require a voltage probing solution that has high CMRR, fast overdrive recovery, voltage clamping (so the oscilloscope is not overdriven), compensation flatness, gain/ amplification to see small signal details, and precise offset generation to see the switching device's turn-off performance. The Teledyne LeCroy **DA1855A** and its associated **DXC Series Passive Differential Probes** are required.

Deskew Signal Source

To assist with eliminating propagation delay differences among voltage and current probes used for device testing, the **DCS015** de skew calibrated source is recommended. This source has time-coincident voltage and current signals used to adjust deskew values within the oscilloscope channel controls and Power Analyzer software.

Method of Operation

The general process for using the software to conduct power analysis is:

1. Set up DUT test circuit, consider isolation requirements, attach probes, and setup oscilloscope trigger

This includes all physical circuit setup and oscilloscope setup for timebase and acquisition triggers. We show recommended connections, probing points, and trigger events for each type of analysis in the preliminary setup topics.

2. Set up voltage and current source channels

Make all Vertical settings on the channels to be used for a test and perform all necessary preliminary adjustments to ensure measurement accuracy.

Because signals associated with power devices are relatively fast, it is important to determine whether the propagation time for the current and voltage signal paths are the same. Signal delay characteristics of the voltage and current probes, as well as the distance the signals must travel from the probe tips to the input of the oscilloscope, can cause time-coincident points on the voltage and current signals to be sampled by the oscilloscope at different times. Even a small time difference can cause significant errors to occur in the measurements. Therefore, we recommend performing a preliminary deskew procedure and repeating it whenever you change to the physical characteristics of the probes or the bandwidth/filter settings of the input channels.

Likewise, Fine DC Adjust voltage channels to remove any residual charge that may be in the probes.

3. Select analysis type, tests to be performed, and measurements to be displayed


This step comprises most of the work you will do in the Power Analyzer Software. Once the source channels are set up, various tests can be performed on the same inputs. The software automatically calculates your selected power measurements and displays them on screen, with test-appropriate trace annotations and overlays to help you see significant portions of the input waveforms.

Power Analysis Software Dialogs

The Power Analysis Software presents a series of dialogs for setting up measurements specific to testing switched-mode power supplies and devices. Generally, the order of the tabs presents the order in which you proceed to use the software, working from left to right.

Power Analysis Dialog

The first tab is the Power Analysis dialog. This is the main set of controls where you select the **Voltage** and **Current Source** input channels, the **Analysis Type**, and the specific **Tests** to run.



The Power Analysis dialog is also where you control the display of Statistics or Histicons within the Power Analysis Measurements table, and Clear Sweeps to reset the measurement value counter.

The **Grid** control allows you to quickly change the grid style. The default setting, PowerAuto, displays the correct number and style of grids for the selected power test. This setting is only available when Power Analysis is enabled.

Quickly return to the Power Analysis dialog from any other dialog by selecting the leftmost section of the Power Analysis measurement table.

Device	TurnOn	Conduction	TurnOff	OffState
value	3.6114017 µJ	1.9872963 µJ	3.9521554 µJ	3.3932410 µJ
mean	3.688251673 µJ	1.766332420 µJ	6.153658590 µJ	-179.951849 nJ
min	3.4602253 µJ	921.8813 nJ	1.9540425 µJ	-4.7064193 nJ
max	4.0348591 µJ	2.6779549 µJ	8.0828663 µJ	3.3932410 µJ
sddev	95.317378 nJ	399.941756 nJ	1.810404735 µJ	1.411018584 µJ
num	654	654	654	436
status				


C1	BvL DC	DevicePower	Current	Voltage
50.0 Vdiv	50.0 Vdiv	500 mA/div	50.0 Vdiv	50.0 Vdiv
160.00 V	2.00 µdiv	2.00 µdiv	2.00 µdiv	2.00 µdiv

Input Settings Dialog

The second tab opens the Input Settings dialog, which allows you to adjust the **Fine DC Offset** and **Deskew** values of your probes to increase measurement accuracy.

You can view the *result* of adjusting Fine DC Offset and Deskew by checking **View** on the Input Settings dialog. This is a convenience to assist with fine adjustment; it's not necessary to keep this trace open. Deskew values are duplicated on the Channel dialog, and the Power Analysis Software incorporates the Fine DC Offset value in its measurement results.

You also use the Input Settings dialog to select the type of device used to measure current in **ProbeType**.




Setup Source buttons on the Input Settings dialog enable you to quickly access the source Channel setup dialogs, where you can adjust input bandwidth limits, set up filters, AutoZero voltage probes, or DeGauss current probes.

NOTE: When selecting Device Analysis, Losses Test and a Conduction Loss calculation method measuring V_{sat} with a 2nd voltage probe, a third set of controls will become available for Deskew and DC Fine Adjust. These controls are made available for use with a probe or amplifier solution that incorporates voltage clamping and fast overdrive recovery, such as the DA1855A/DXC100A.

Zone Identification Dialog

This tab appears only when the Analysis Type is Device. The Zone Identification dialog is used to set up device switching measurement zones, which adds a set of color overlays and annotations to the Power Analysis trace.



Other Dialogs

Tabs for other dialogs, such as Device Power, appear only when the corresponding Analysis Type and Test are selected. They contain rescale controls that allow you to “zoom” the result trace to view more or less waveform detail. The controls are the same as found on any Zoom dialog, although in this case they will alter the appearance of the Power Analysis trace instead of opening a new zoom trace.

Deskew Voltage and Current Channels

Use this Deskew procedure to check propagation delay differences between the current and voltage channels. This is very important if these signals are going to be used to make instantaneous power Losses, Safe Operating Area, or Dynamic On-resistance measurements.

This process can be used to characterize and correct the delay difference between more than one current and one voltage channel. For instance, if you plan to use one voltage channel and alternate measurements between two current channels, the relationship between the voltage channel and each of the current channels can be characterized. The same is true if you are using a differential amplifier to capture voltage at key event points. The amount of deskew required for each combination should be recorded for later use.


Throughout this procedure, the Voltage source channel is used as the reference trace. Adjust all other traces to this reference.

1. Recall the oscilloscope's factory default settings.
2. Connect the:
 - Voltage probe to Channel 1
 - Current probe to Channel 2
 - DCS015 Deskew Calibration Source to the EXT input.
3. Set **Channel 1 scale to 1 V/div**. If using a differential probe on Channel 1, also **AutoZero** the probe.
4. Set **Channel 2 scale to 20 mA/div**. **DeGauss** the current probe.

NOTE: It's good practice to leave the current probe disconnected from the test circuit due to the excitation signal used to DeGauss the probe.

5. Connect the voltage and current probes to the DCS015. Be sure to match the proper voltage probe polarization and current probe direction to the DCS015.
6. Set the oscilloscope **timebase to 20 ns/div** and zero delay.


7. Set the oscilloscope **trigger to Channel 1, negative edge, at a level of 2 Volts.**
8. Choose **Display > Single Grid**, then **adjust Offset** on Channel 1 and Channel 2 so that the voltage and current waveforms are on top on each other in the middle of the display.



9. Choose **Analysis > Power Analysis** to open the Power Analysis Software.
10. On the Power Analysis dialog, select C1 as the **Voltage source**, and C2 as the **Current source**.
11. On the Input Settings dialog, adjust the Current **Deskew** value until the slope of the current probe intersects the voltage waveform at the upper knee of the falling edge.

Tip: You can do this by selecting the Deskew field, then turning the Front Panel Horizontal knob.
12. Repeat the deskew procedure for the differential amplifier or any other probes you have connected to other channels.

NOTE: The Deskew values you obtain using this procedure are only valid for this particular setup. It is recommended to repeat the deskew procedure if you change probes, cables, or bandwidth/filter settings on any channel.



Falling edge of voltage and current traces before deskew adjustment.



Falling edge of voltage and current traces after deskew adjustment.

Power Device Analysis

Device Analysis Preliminary Setup

The Device Analysis tests let you make difficult measurements on devices while they operate in circuit. The exact setup for each measurement will differ depending on what device type is to be analyzed and where it is located in the circuit.


Test Circuit Setup

Examples in this section are based on connections to an off-line flyback power supply circuit. Measurements are made on devices such as power transistors, snubber diodes, or similar devices in other topologies.

A typical setup used to analyze the power MOS-FET in an off-line switching power supply is:


- A differential high-voltage probe is connected to V_{ds} on the oscilloscope's Channel 1.
- A current probe is connected to the drain current, I_d , using a current loop into Channel 2.
- A differential amplifier, with a matched differential probe pair, is used to connect to either V_{ds} or V_{gs} into Channel 3. This amplifier will need to have voltage clamping and fast over drive recovery in order to see the saturation voltage and have high CMRR to capture the high-side gates in an off-line application.

The example in the figure below uses the oscilloscope's Channel 4 to acquire a trigger signal indicating when the load changes from maximum to minimum. You could also use the oscilloscope's EXT trigger input.




Trigger Setup

Triggers for Device analysis are generally associated with a load change or on/off event in the circuit.



Obtain a stable display of at least one cycle of the voltage signal across the device.



Set Up Device Analysis Input Channels

This procedure selects the voltage and current source channels and makes any necessary offset or deskew adjustments to ensure measurement accuracy.

If you have just completed the preliminary deskew procedure, begin at Step 6.

1. From the menu bar, choose **Analysis > Power Analysis** to open the Power Analysis dialog.
2. Check **Enable**.
3. Select the **Voltage Source** channel where the voltage probe is attached.
4. Select the **Current Source** channel where the current probe is attached.

NOTE: The current input channel is automatically assigned Ampere units.

5. Touch the **Input Settings** tab and enter the current probe **Deskew** value obtained during the preliminary deskew procedure.
6. Choose the current **ProbeType**. If you're using a current device other than a probe (e.g., shunt resistor), also make these entries:
 - **If an amplifier**, also enter the **Amps/div**.
 - **If a shunt resistor**, also enter Shunt Resistor **Ohms**.
7. If necessary, touch **Select Source** and make any final adjustments on the voltage or current Channel setup dialogs.


NOTE: If you change the probe, bandwidth or filter settings at this point, repeat the [preliminary deskew procedure](#) and re-adjust the Deskew setting on the Input Settings dialog.

8. Choose **Display > Auto Grid** and set up peak-to-peak waveforms to utilize the full Analog-Digital Converter (ADC). The waveform should cover the full height of the grid.
9. When you have obtained a stable display, go on to [identify the switching zone](#).

Identify Switching Zone

This procedure instructs the software how to recognize the transition event in the device power trace (e.g., ON to OFF) and focus the analysis on it. To do this, you'll define the vertical area of interest (Signal Top and Base) and the horizontal area of interest (Transition Threshold) as a measure of voltage.

1. [Set up the source channels.](#)
2. Touch the **Zone Identification tab** and choose a **Top Base Value** of either:
 - **TopBase**—analyze the entire waveform top (mean of all max. V) to base (mean of all min. V)
 - **Levels**—analyze only voltages between this absolute Top Level and Base Level



3. Choose whether the transition **Thresholds Are** given as a **Percent** of signal or **Absolute** values, and enter the **High Threshold** and **Low Threshold** representing the starting voltage and ending voltage that marks the transition event in the waveform.

Overlays appear on the voltage trace marking the Conduction, On, and Off states, similar to this:



Go on to perform your desired Device Analysis tests.

[Losses Test](#)

[Dynamic On-resistance Test](#)

[Safe Operating Area Test](#)

[BH Curve Test](#)

[Device dv/dt Test](#)

Device Losses Test

This test measures switching device power losses as:

- Mean power—Watts
- Peak power—instantaneous peaks in Watts
- Energy—Watts/seconds in Joules

1. [Select the source channels](#) and [identify the switching zone](#).
2. On the Power Analysis tab choose **Analysis Type Device** and **Test Losses**.

3. Select the **Measurements** to calculate. Options are:

- Switching losses – losses from switch between states
- Conduction loss—loss from internal resistance
- Offstate losses—losses during the off state due to instrumentation as noted below

NOTE: Offstate loss in discontinuous power devices is typically zero, and any value is probably due to residual DC in the current probe. Therefore, this measurement is useful for fine adjusting the current probe. Adjust the Fine DC Offset field of current probe input settings until the measured Offstate loss value is as close to zero as possible (ideally in Nano Joules).

- Sum—total measured losses over entire switching cycle
4. Choose to display the measured **Mean Power**, **Peak Power**, or **Energy (Joules)**.

NOTE: Selecting "Measure Over All Selected Zones" uses all of the zones selected as the region for the measurements. It will not sum each of the zones individually to give a final result. This is especially true with displays of Mean Power and Peak Power.

5. If you have selected to calculate Conduction Loss, choose the conduction **Calculation Method**.

Options are:

- **Std. Probe Inputs**—Select if you are using only the standard two probes.
- **2nd Voltage Probe**—Select if you are using a differential amplifier (or 2nd voltage probe) to measure saturation voltage at the event site. Select the Voltage Source channel where the amplifier is connected.
- **MOSFET Rds(on) Value**—Best if your circuit is a MOSFET and you know the Resistance of the conductor. Enter the value in Ω .
- **IGBT Vce(sat) Value**—Best if your circuit is an IGBT and you know the Saturation Voltage. Enter the value in V.

Readouts for the selected measurements appear in a table below the grid area.



6. Optionally, select **Statistics** to display mean, minimum and maximum values on the measurements table. This helps to observe a consistent number.
7. Optionally, select **Histograms** to display a miniature histogram of the statistical measurements.

Device Safe Operating Area Test

This test measures that voltage, current, and power of an event all fall within safe operating levels as defined by the device manufacturer. You can apply a mask to the power trace to help set your operating limits or operating margin expectation for the device.

1. [Select the source channels](#) and [identify the switching zone](#).
2. On the Power Analysis tab choose **Analysis Type Device** and **Test Safe Operating Area**.

The drain-to-source voltage waveform is displayed on the first grid, and the deskewed current waveform is displayed on the second grid.



On the XY plot in the third grid, voltage points are plotted on the x-axis, while current is plotted on the y-axis. Therefore, the bottom left corner of the XY trace is the 0,0 point; the top of the trace is maximum current; far right of the trace is maximum voltage. The most frequently occurring samples are marked by a higher intensity display.

3. To apply a mask test to the power display:
 - Check **Mask Test**.
 - Enter the bounds for the mask by entering the device limits defined by the manufacturer or your desired test limits. The value represented by each grid division appears on the channel trace descriptor boxes.

A circular red marker appears around each data point that exceeds the mask limits.
4. Optionally, select **Statistics** to display mean, minimum and maximum values on the measurements table. This helps to observe a consistent number.
5. Optionally, select **Histograms** to display a miniature histogram of the statistical measurements.


Device B-H Curve Test

This test displays the B-H hysteresis plot that confirms the saturation status of a magnetic device. The voltage across a waveform, typically acquired with a differential voltage probe, is set as the voltage source. The current through the device is captured typically with a current probe. The hysteresis plot is presented as the integrated voltage across the magnetic device versus the current through the device.

1. [Select the source channels](#) and [identify the switching zone](#).
2. On the Power Analysis tab choose **Analysis Type Device** and **Test B-H Curve**.

The scope display shows the voltage source before it is integrated, the current source, and the hysteresis plot.

3. Select to **View Magnetic Flux Density (B)** and/or **View Magnetic Field Strength (H)**. Additional result traces appear for each selection.



4. Optionally, enter the Coil Characteristics in **# Windings**, **Cross Section Area** and **Magnetic Path Length**. This will set the appropriate scale values within each of the descriptors. Cursors provide additional details of the values on specific parts of the B and H waveforms.
5. Optionally, select **Statistics** to display mean, minimum and maximum values on the measurements table. This helps to observe a consistent number.
6. Optionally, select **Histograms** to display a miniature histogram of the statistical measurements.

Device Dynamic On-resistance Test

This test measures the on-state resistance of a switching device. It utilizes the saturation voltage and current measured during the on-state zone and automatically calculates the $R_{ds(on)}$ value, displayed in the parameter table. For off-line power supplies, it is recommended to use a differential amplifier, such as the DA1855A and DXC100A, in order to clamp the output voltage level the oscilloscope receives when over-driving to see saturation voltage details.


1. [Select the source channels](#) and [identify the switching zone](#).
2. On the Power Analysis tab choose **Analysis Type Device** and **Test Rds(on)**.

The deskewed voltage (V) and current (I) waveforms are displayed on the first grid.

3. Adjust voltage trace **Scale (V/div)** and **trigger Level** until the voltage waveform displayed is the device saturation voltage.

The voltage waveform is divided by the deskewed current waveform, and the resulting Resistance waveform trace is displayed on the third grid.

NOTE: Because the device saturation voltage waveform is off screen during the non-saturation portion of circuit, the math-generated resistance waveform should be ignored during this time.



4. Optionally, select **Statistics** to display mean, minimum and maximum values on the measurements table. This helps to observe a consistent number.
5. Optionally, select **Histograms** to display a miniature histogram of the statistical measurements.

Device dv/dt Test

This test measures the speed of a power device's rate of change (dv/dt) during turn-on and turn-off.

1. [Select the source channels](#) and [identify the switching zone](#).
2. On the Power Analysis tab choose **Analysis Type Device** and **Test dv/dt & di/dt**.
3. Select di/dt by checking the **di/dt** box in the Waveform section.

The power device's drain-to-source voltage (VDS) signal and the drain current are displayed on the first grid. The derivative of the waveform's dv/dt is displayed on the second grid. The di/dt waveform is shown in the third grid.




4. Optionally, enter an **Enhance Bit Value** in the Setup section. This applies a filter to the data to reduce noise.
5. Optionally, select **Statistics** to display mean, minimum and maximum values on the measurements table. This helps to observe a consistent number.
6. Optionally, select **Histicons** to display a miniature histogram of the statistical measurements.

Control Loop Analysis

Control Loop Analysis Preliminary Setup

Switched-mode power conversion circuits use some method of transferring energy from an unregulated source to regulated outputs on a cycle-by-cycle basis. Output regulation is achieved by modulating the amount of energy transferred in each cycle, the most common method being Pulse Width Modulation (PWM). Other methods (such as frequency modulation) are also used.

Control Loop analysis provides you with tools to view the information contained in the control circuit's modulated signals. It does this by taking the time (duty cycle or width) information in the modulated signal that is normally displayed on the horizontal axis, along with elapsed time, and displays it on the vertical axis.




As the number of pulses per division increases, the display of their individual widths forms a “waveform” that represents the change in pulse width (or other characteristic) as a function of elapsed time. This “waveform” can be used to gain valuable information about the power supply’s response to various events, such as load change (step response) or its soft-start performance.

Test Circuit Setup

Control Loop Analysis lets you capture and analyze information contained in the power conversion circuit's modulation. The exact setup for this measurement will differ depending on the specific circuit topology and where in the circuit under test the modulation signal is to be acquired.


Examples in this section are based on a typical setup to acquire the modulated signal at the output of a Controller IC, as shown in the image below.



For off-line power supplies, the Teledyne LeCroy DA1855A Differential Amplifier is used to acquire the device's gate drive signal.

Trigger Setup

Control Loop analysis measurements usually are made to find the circuit's response to some event. Turn-on, turn-off, line trigger, or load change can be used to trigger the modulated signal acquisition. In the example below, a load change on the output is used as the event trigger, and the power transistor's gate drive signal is used as the source of modulation information.



When setting up the Control Loop trigger:

- Set up a stable display of the signal that will be used as the source of the modulation information. Ensure that a clean signal can be acquired that will allow the signal's width (or other characteristic) to be readily measured.
- If the modulated signal is to be acquired as the result of a one-time event such as turn-on, test the trigger for satisfactory operation in Single trigger mode. If the event is repetitive, you can use the oscilloscope's Normal trigger.
- Be sure to set trigger Delay and Time/div to a value that will allow the capture of a modulated signal record sufficiently long to cover the time of interest.

Set Up Control Loop Analysis Source Channel

This procedure selects the voltage source channel and makes any necessary offset or deskew adjustments to ensure measurement accuracy.

If you have just performed the preliminary deskew procedure, begin at Step 5.

1. From the menu bar, choose **Analysis > Power Analysis** to open the Power Analysis dialog.
2. Check **Enable**.
3. Select the **Voltage Source** channel where the voltage probe is attached.
4. If you are using a current device to trigger on a load change event:
 - Touch the Input Settings tab and enter the **Deskew** value obtained during the preliminary deskew procedure.
 - Choose **ProbeType**.
 - If the current device is other than a probe (e.g., shunt resistor), also make these entries:
 - If an **amplifier**, also enter the **Amps/div**.
 - If a **shunt resistor**, also enter Shunt Resistor **Ohms**.
5. Adjust **Fine DC Offset** to zero the level.
6. If necessary, touch **Select Source** and make any final adjustments on the voltage or current Channel setup dialogs.

NOTE: If you change the probe, bandwidth or filter settings at this point, repeat the [preliminary deskew procedure](#) and re-adjust the Deskew setting on the Input Settings dialog.

Obtain a stable display of at least one cycle of the voltage signal across the device, then go on to perform your desired tests.

[Closed Loop Test](#)

Closed Loop Test


1. [Set up the source channel.](#)
2. On the Power Analysis tab choose **Analysis Type Control Loop** and **Test Closed Loop**.
3. Choose the **Operator** (method) used to modulate the signal. This is the input signal parameter that will be used to generate the Closed Loop trace.
4. Choose the **Slope** (Pos. or Neg.) on which to measure.
5. Touch the **Level Is** field and select to enter the measurement voltage Level as an Absolute Value, Percentage (of top to base), or one of the percent ranges.
6. Enter the absolute or percentage of **voltage Level** at which to measure the operator. Set this to a level on the modulated signal where both the rising and falling edges are free of noise.

Tip: When measuring the modulation of the gate drive signal, it is best to avoid placing the level around the pedestal.

OR

Touch the **Find Level button** to let the software set Level based on the amplitude of the acquired gate signal.

The gate-drive signal is captured and displayed in the top trace. With the load change as the trigger event, the controller's response is displayed as the blue trace in the second grid. Drag-and-drop the Closed Loop descriptor box on any other grid to move the trace.




Line Power Analysis

Line Power Analysis Preliminary Setup

The Line Power Analysis section of PMA2 provides you with tools to measure 50 and 60 Hz line voltage (Vrms), line current Irms), Apparent Power (VA), Real Power (Watts), and Power Factor ($\cos \theta$). It also enables you to evaluate harmonic currents injected into the power line. Harmonic measurements are made in accordance with the requirements of standard EN 61000-3-2, Harmonic Current Emissions. Teledyne LeCroy encourages you to refer to the latest version of EN 61000-3-2 for full definitions and limits set forth by the standard. The information here is abstracted for your convenience.

Test Circuit Setup

In the examples in this section, Channel 2 is used for voltage and Channel 3 is used for current, but any channel could be used for voltage or current. In the case of 3-phase systems, multiple voltage and/or current channels can be set up before analysis is started.



The current waveform polarity must match that of the voltage waveform. Make sure your current probe is connected so that this is the case.

Trigger Setup

Triggers for Line Power tests can be either the voltage or current waveforms.

Set Up Line Power Analysis Source Channels

This procedure selects the voltage and current source channels and makes any necessary offset or deskew adjustments to ensure measurement accuracy.

If you have just completed the preliminary deskew procedure, begin at Step 6.

Power Analysis Software

1. From the menu bar, choose **Analysis > Power Analysis** to open the Power Analysis dialog.
2. Check **Enable**.
3. Select the **Voltage Source** channel where the voltage probe is attached.
4. Select the **Current Source** channel where the current probe is attached.

NOTE: The current input channel will be assigned Ampere units even if a voltage probe or a non-Probus probe is used.

5. Touch the **Input Settings** tab and enter the **Deskew** value obtained during the preliminary deskew procedure.
6. Choose the **ProbeType**. If you're using a current device other than a probe (e.g., shunt resistor), also make these entries:
 - **If an amplifier**, enter the **Amps/div**.
 - **If a shunt resistor**, enter Shunt Resistor **Ohms**.
7. Adjust **Fine DC Offset** to zero the level.
8. If necessary, touch **Select Source** and make any final adjustments on the voltage or current Channel setup dialogs.

NOTE: If you change the probe, bandwidth or filter settings at this point, repeat the [preliminary deskew procedure](#) and re-adjust the Deskew setting on the Input Settings dialog.

Obtain a stable display of at least one cycle of the voltage signal across the device, then go on to perform your desired tests.

[Line Power Test](#)

[Line Harmonics Test](#)

Line Power Test

This test analyzes the line current, voltage, power, and power factor of the device under test.

1. [Set up the source channels.](#)
2. On the Power Analysis tab choose **Line Power Analysis**, and **Power Test**
3. Optionally, select to display **Reactive Power**, **Phase Angle** and/or **Crest Factor** measurements.


The display shows the line voltage and current waveform, as well as the power and energy waveforms. Parameters displayed are rms voltage, rms current, real power, apparent power (apwr) power factor (pf) and, if selected, reactive power (reactpwr), phase angle (phase), and crest factor (Icrest).



Line Harmonics Test

This test checks the line harmonics to see if they meet EN 6000-3-2 pre-compliance levels. In addition, for class C, tests differentiate devices above and below 25 W; for class D, a Current per Watt pre-compliance test is provided.

1. [Set up the source channels.](#)
2. On the Power Analysis dialog choose **Line Power Analysis** and **Harmonics Test**.
3. Touch **Class** and select the equipment's EN 61000-3-2 classification from the pop-up menu:
 - **Class A**, Balanced three-phase equipment, household appliances (excluding those in class D), non-portable tools, audio equipment, dimmers for incandescent lamps, and all other equipment except that stated in one of the following classes
 - **Class B**, Portable tools, non-professional arc welding equipment
 - **Class C**, Lighting equipment (except dimmers)
 - **Class D**, Personal computers and computer monitors, radio or TV receivers (75W-600W)
4. If you selected Class C, open the Limits dialog and choose the correct **Active Input Power** and **Equipment Type**.



5. On the Power Analysis dialog, touch **Frequency** and select the line frequency at which the equipment is operating. You can optionally select to **Use Measured Frequency** to enter the value as measured by the current probe.
6. Touch **Units** and select Amps or dBuA (decibels referred to one microAmp).
7. Optionally, select **Total Harmonic Distortion** to see the contribution of each frequency to the total distortion.



TIP: You may need to open the Line Harmonics or Current per Watt dialogs to adjust the zoom of the trace.

To see the harmonics readout in the measurements table:

- Stop acquisition by pressing the front panel **Stop trigger** button.
- Choose **Cursor > Horizontal Abs** from the menu bar.
- Turn the front panel **Cursor knob** or touch-and-drag the white cursor line to move it along the line harmonics trace.

Performance Analysis

This section of the Power Analyzer option provides tools to evaluate the performance of your power supply.

- The Efficiency Test lets you measure input power, output power, and overall efficiency.
- The Ripple Test measures ripple—including pk-pk, maximum, and minimum ripple—and generates a spectral plot of the frequency content of the power supply's output.


Because of the varying ranges of voltages and currents of different power supply topologies, including requirements to maintain isolation, please use the appropriate probes when performing these tests. Be sure to maintain the proper probe setups and make fine adjustments to DC offsets and DeGaussing using the Input Settings tab controls.

Performance Efficiency Test

This test checks the efficiency of the power conversion by computing power on output divided by power at input.

Efficiency Test Setup

For input power, attach the voltage and current probes to the corresponding input sources shown below. Output Power is measured using the voltage and current probes as shown below. Confirm the Power Analyzer software input and output probe sources are matched to this setup.



Efficiency Test Procedure

1. From the menu bar, choose **Analysis > Power Analysis**.
2. Check **Enable**.
3. On the Power Analysis dialog choose **Performance Analysis** and **Efficiency Test**.
4. Choose the **Input Voltage**, **Input Current**, **Output Voltage** and **Output Current** channels.
5. Optionally, select **Statistics** to display mean, minimum and maximum values on the measurements table. This helps to observe a consistent number.
6. Optionally, select **Histograms** to display a miniature histogram of the statistical measurements.




Performance Ripple Test

This test supports the measurement and analysis of a power supply's output characteristics. Measurements include the peak to peak ripple, the maximum and minimum values of ripple, and the spectral content of ripple are provided.

Ripple Test Setup

Sources for this section can be voltage or current. Confirm the correct input source has been selected in the main Power Analyzer menu. Connect the probe to the output of the power supply.



Ripple Test Procedure

1. From the menu bar, choose **Analysis > Power Analysis** to open the Power Analysis dialog.
2. Check **Enable**.
3. Choose **Performance Analysis** and **Efficiency Test**.
4. Select the input **Voltage** source channel.

5. On the Input Settings dialog, touch **Setup Source** and choose **AC Coupling**.





700 Chestnut Ridge Road
Chestnut Ridge, NY 10977
USA

teledynelecroy.com